

Medical Services Department

Specifications:

Dental Compressor:	<p>Safety & Control Components: Pressure Regulator, Pressure Gauge, Safety Valve & Non-return Valve.</p> <p>Type: Oil Free Working Pressure: 5 – 8 bar Motor Power: 750 – 850w Voltage: 220v Air Tank Capacity: 30 - 50 Noise Level: 60 – 70 dB (silent type)</p>
---------------------------	--


Dr.amna Urooj Siddiqui
B.D
Medical Services Department
Sui Southern Gas Co. Ltd

Medical Services Department

Specifications:

<p>Patient Couches:</p> <p>CC136831</p>	<p>Stainless steel frame: Standard sizes range from Height <i>length</i> 5.8 to 6 ft Width: 3 to 4 ft wide. <i>2.5 to 2.6 ft wide</i> ZAB</p> <p>Foam (2-3 inches thick) covered in Rexene</p> <p>Standard load capacities are typically: 100--175 KG</p>
<p>Tens Machine:</p> <p>CC137221</p>	<p>Power Supply: DC 9V Number of Channels: 2 Safety Class according to IEC 60601-1 Internally powered equipment. Type BF Output mode: Constant, Burst, Modulation Frequency: 200Hz</p>
<p>Ultrasonic Machine:</p> <p>CC136705</p>	<p>Operating Conditions: Ambient temperature : -10cJ ~ 40cJ relative humidity: below 80%</p> <p>Power Supply : 220 V</p> <p>Output Power: Continuous wave: 0.53 W/cm² (adjustable in 11 steps)</p> <p>Pulse wave: 0.15 – 1 w/cm²</p> <p>Operating frequency : 800khz i0 5%</p> <p>Power Consumption: Max. 70w</p> <p>Probe: Calcium Zirconate titanate, non-water cooling type two transducers included</p>

[Signature]
 6/3/26
 Dr. Rooman
 Shah